Technical Specification of Operation Theatre Positioning Devices



Supine or Contoured Head Rest Standard Green No Film

	Generic / Technical	Compliance Statement(s)
Name of Item	Specifications Supine or Contoured Head Rest	Supine or Contoured Head Rest Standard,
Nume of item	Standard, Green No Film	Green No Film
Technology/ Product Specifi	cations	
Product Category	As mentioned in catalogue or	Supine or Contoured Head Rest(s)
Purpose of Product	website: otcare.co.in As defined in catalogue	 Relieves skull compression from head/
Fulpose of Floudet	As defined in catalogue	calf in supine position (like eye surgeries)
		by providing cross-sectional stability
		 Contoured to match the anatomical shape of head by pressure redistribution in all
		directions
Product Design (with	Product Image with marking of	
markings)	dimensions	Inner Diameter (d)
		Median Median Height (h)
		Median Height (h) Length (L)
		rengen
Material	Medical Grade cured gel which	Yes
Type of Gel Technology	doesn't flow out Viscoelastic Gel with Covid	C3I Gel
Type of Ger resimology	Coat+ (Anti-microbial coating) or	
Comparis To also also an Name	C3I Gel	031.0-1
Generic Technology Name Brand Name or OEM	C3I Gel (CC+) or C3I Gel Whether the item is required	C3I Gel Without Brand Name: OEM or optional as per
Drana namo en e Em	with/ without Brand Name i.e.	buyer
Color of Droduot	OEM	Croon Emorald
Color of Product	Sapphire Blue/ Emerald Green/Custom Color	Green Emerald
Type of Construction	Film/ No Film (single continuous	No Film (Single Continuous Phase)
Charifications of Home Marie	phase)	
Specifications of Item /Varia Item Code	As per listed SKU Code	EGR-NF-SHR - SD
Size of Item, Number of Items	Neonatal/ Small/ Medium/	
Per Pack	Adolescent/ Large/ Adult/ Extra	Standard (SD), Piece Per Pack
Dimensions of Item	Large/ Standard Size As mentioned in catalogue	L*W*H [d*h] (in mm) = 175 X 80 X 70 [145 X
- Dimonsions of item	7.5 Montioned in Catalogue	2 17 17 [W 11] (11 11 11 11) 17 0 17 0 0 17 10 [1 1 1 0 1 1

each annotation used. Height; d = Inner Diameter; h
Dimensional marking of item


Grey Image of Product Item with

Grey Image of Product Item with specific marking of each dimension

specific to item with definition of


35] mm, wherein L= Length; W= Width; H= Height; d = Inner Diameter; h = Median Height





Dimension of Pack (Wooden: Cardboard Box)

Dimension of Wooden: Cardboard Box (Length X Breadth X Height) from exterior side (in cms)



Box Number 12: 28.8 X 24 X 10 cms

Dimensional Marking of Pack (Wooden: Cardboard Box)

Marking of Exterior Length, Breadth and Height of Wooden: Cardboard Box (in cms)



Net weight of Item

Gross Weight (Item with Wooden: Cardboard Box)

Weight of item (in grams) without any accessories or packing Gross weight inclusive of item, its accessories and Wooden:

Cardboard box used for packing (in grams)

Specification (with optional choices)

Pandemic Resilience Ability: Prevent Spread of Infection

Specialized Gel Technological
Variant with Covid Coat+
(sustainable solution that is
tested against hospital

pathogens responsible for

No

2542 grams

Dookoaina, Internel	Hospital Associated Infections)	
Packaging: Internal	Option 1: Viscoelastic Gel with	
Environment assurance and	Covid Coat+ (with/ without) film	No accessory items are provided in the box.
accessories inside pack	based item should have anti-	
	microbial sticker, air clean pouch	
	and 50 ml of Covid Coat+	
	solution.	
	Option 2: C3I Gel (with/ without	
	film) doesn't contain any	
O - man l'anne a Chatana ant a l'illa	accessory items	
Compliance Statement with	Option 1: Compliance	No seeds a seed of seeds information money in a
AST SOPS (Digital Access)	Statement (Generic)/ Specific to	No such access of such information provided
(Optional)	Product should be provided in	in the pack.
	physical or digital format	
	ensuring patient safety as per	
	laid down Guidelines by	
	Association of Surgical	
	Technologist(s) SOPs for	
	Positioning Devices Option 2: No such access of	
	information provided with pack	
Barcodes (optional)	Option 1: Bar codes specifying	No barcode specifying company information is
barcodes (optional)	item information should be	provided on the box.
	available outside box.	provided on the box.
	Option 2: No Barcode is	
	provided on the pack.	
Acceptance Test Procedures	Option 1: User Manual provides	
Compliance (optional)	description of ATPs which	User Manual is not provided with Testing Kit in
,	minimally suffice testing	this pack.
	parameters to establish (a)	
	environmental friendly material;	
	(b) elastomerism; (c)	
	hydrophobicity; (d) cold	
	tolerance; (e) impact resistance	
	and; (f) destructive testing,	
	thereby ensuring the nature of	
	gel technology to support the	
	requirements of operation	
	theatre positioning device	
	Option 2: User Manual is not	
	provided with testing Kit	
User/ Technical /	Option 1: Bilingual (in physical	
Maintenance manuals to be	and digital copy) User Manual	No User Manual is provided with Packaging
supplied in English & Hindi in	with third optional language	Kit.
hard and soft Copy	(digital access) as per specific	
	provision of order only (if	
	provided) should be available.	
	Option 2: No User Manual is	
	provided with Packaging Kit.	



Manufacturing Certifications and Product Standards Operation Theatre Positioning Devices

	Generic / Technical Specifications	Compliance Statement(s)
Product Standard Certification	CE or any other standard (if applicable) Certification Number, Date	CE Certification Number: CE-97801 Date: 09-04-2021 to 08-04-2024
Manufacturer's Quality Standard	ISO 9001: 2015 and ISO 13485: 2016 as well as any other standard with its number and date	ISO 9001: 2015 Cert No : 305021041024Q Date: 10-04-2021
		ISO 13485: 2016 Cert No: MD-9014121 Date: 09-04-2021
Certification regarding chemical inertness	REACH standard assuring Gel free from any hazardous /environment degrading constituents Cert Number & Date	REACH Cert No: CH:TX: 1142028342 Date: 29-09-2020
GMP Certification	Standards for Global Manufacturing Practices- Best Practices adopted for development of medical devices i.e. WHO: GMP Cert No and Date	WHO: GMP Cert No: UQ-901241 Date: 09-03-2021 to 08-03-2024
QCI Certification	Certification issued by Quality Council of India (QCI) for adoption of best quality practices in development, occupational hygiene, management, safety and other aspects.	QCI Cert No: VA-452406728758919 Date: 27-7-2021 to 26-07-2024



Technical Specifications: Material, Technology, Packaging, Purchase & Site Utilities

Operation Theatre Positioning Devices

Category	Generic / Technical Specifications Health Care Device/ Non-Health Care Utility Device/ Both	Compliance Statement(s) Health Care Device
Toxin Free Composition	Material to be Latex free Material to be Phalates free Material to be Polyurethane Free Plasticizer & silicone free Gel	Latex, Plasticizer, Phalates, Silicone and Polyurethane Free Gel.
Human Safe	Material to be non toxic, Skin compatible and inert	Compatible with skin even during long duration usage while in direct contact with skin during surgeries
Eco-friendly and Recyclable	Material shall be eco-friendly and recyclable	Gel material is REACH certified, thus environment friendly as well can be recycled.
Elasticity of Material (Elastomeric Cushioning)	Item is designed to evenly distribute the body pressure to prevent circulatory disturbance and pressure ulcer at the bony prominences	Elastomeric Cushioning Gel complies with minimum requirement of Elongation strain withstands ability (≥ 500%) while simultaneously Shore A value limits to Zero.
Elongation Strain Ability (%)	Elastomeric Ability is estimated as elongation strain withstands ability with minimum of 500%.	543%
Soft Cushioning	Material having Shore A value limits to zero	Item's material has Shore A value limits to Zero indicating its extreme softness preventing pressure sore
Conforming Ability (Complete/ Partial)	Device should be able to complement body's contours completely	Complete: Designed according to body contours to prevent intra operative skin injuries and to avoid respiratory or nerve injuries
Non-Leaky	Material should not flow on prick	Elastomeric Gel is closed cell structure that doesn't leak out when pricked or punctured.
Density	0.3 to 0.9 g/cc "Optimal value attributes towards strength of Positioning Device to hold body weight while not losing flexibility"	Density of material used in development of item has density of 0.8 g/cc (4.9 lb/foot).
Compression Set in	Should be least so that it does not get	3.3

0	. ,	$\overline{}$
-11	- (
- 10	U	\cup

Material (%)	compressed with time.	
Repeated Usage : Bottom	Device should be reusable and	Organic, inert & reusable which don't bottom
Out Resistance Ability	should not bottom out on repeat	out on repeated usage due to high density
	usage	with least compression set.
Weight bearing capacity	≥200 Kg	Item is able to hold 200 Kg of patient's
of the gel pad/ device (in		weight at vertical angle of placement.
Kg)		
Rebounding Capacity	Device should not be immersive and	Item has high rebounding capacity, on
	rebound in less than 1 second	withdrawal of external pressure, in less than
		fraction of second it rebounds, thereby,
		preventing pressure due to any bony
		prominence.
Intrinsic Impact Resistant	Impact Resistant ability is linked to its	Gel material used in development of item
Ability	tensile strength with optimal core	complies with requirement of optimal core
	value of 10 Kg/ cm³.	value of Tensile strength of 10 Kg/ cm ³ .
Tensile Strength of the	Optimal core value of 10 Kg/cm ³	11.4 Kg/ cm ³
technology (Kg/ cm³)		
Crush Strength	No permanent deformation even after	Yes, material complies with such
	n = 1000 crushing cycles at 1600 psi.	requirement for Crush Strength.
Uniform Pressure	Complete / Partial	Complete pressure re-distribution realized as
Distribution		tear strength ranges from 1.0 -4.5 kN/m and
		also reduces capillary interface pressure
		thereby preventing bed sores, pressure,
		shear and frictional impact during prolonged
To ar Ctrongth (I/N/m)	Chould be in rouge of 1 A F I/N/m	surgeries
Tear Strength (kN/m)	Should be in range of 1-4.5 kN/m	4.17
Capillary Interface Pressure Reduction	Able to reduce capillary interface	Yes, it complied with this requirement.
	pressure to 32 mm Hg or less	
Ability V. Day Observation	Davisa should be translugant to V	Itom is translugant to V roy
X- Ray Observation Compatibility	Device should be translucent to X-	Item is translucent to X-ray
Non-Conductive Material	ray. Device should be non-conductive	Itom is non-conductive, an assential preparty
INOTI-CONQUCTIVE Material	Device Should be Horr-conductive	Item is non-conductive, an essential property for its utility in magnetic environment as well
		as doesn't produce static charge.
C ARM / MRI Compatible	Device should be compatible for	Item is compatible for usage in C ARM and
	usage in C Arm and MRI	MRI sections of hospital due to its non-
	asage in 67 mill and wird	conductivity.
		Conductivity.
Range of the usage	12-40°C	Item is maximally effective in this ambient
temperature(in ^o Celsius)	12 10 0	range of 12-40°C.
	The material chould not malt harden	
Cold Crack Resistance	The material should not melt, harden,	Gel Material used in this item is able to

	freeze or crack under standard operating conditions and withstand upto -30°C during storage / transit	withstand up to -30°C with no cracks.	
Non-supportive to microbial growth	Hypoallergenic/ non-toxic & doesn't support any microbial growth	Gel material by its intrinsic composition doesn't support any microbial growth.	
Anti-Microbial Coating	Standard anti-microbial coating is required as per Infection Control Guidelines of Hospital.	Item is coated with anti-microbial coating to prevent growth of opportunistic pathogens	
Washability	Positioning Device can easily be cleaned using standard operating room/ hospital based detergents.	Easily cleaned using Standard Hospital Detergents, instruction manual is provided in every pack as well as available digitally with QR code on each pack.	
Hot Operation Theatre Compatible	Positioning Device that can be used directly with partially contaminated victim of Chemical and Radiological Emergency for Emergency Surgeries in Hot OTs	Compatible for decontamination procedures to remove C& R agents being hydrophobic with closed cell structure (moisture resistant)	
Hydrophobic	The material should be Hydrophobic	Material dose not absorb any water, chemical or fragrance etc	
Leakage proof	The material should be leakage proof	Material, being closed cell structure, is leakage proof & moisture resistant	
Unusual Movement Resistance of OT Table: Adhesion of Material	Velcro to be provided for adhesion.	Device is provided with Stick on Surface that can be attached to Velcro to affix device on OT Table preventing unwanted movement of device during Prolonged Surgeries.	
	Packaging Mater	ial	
Complimentarily of Pack Size, Material, Strength with Item	Suitable packing to withstand transit and accommodate the size of device to be provided	Wooden Box complement to weight of item with size dimensions for ease of fitting with least possible movement during transit	
The type of the inner wall of the packing box	The inner walls are laminated coating to prevent any direct contact with absorbent material or if lamination is restricted in region specific regulation on plastic, butter paper is used.	Inner walls of Wooden Box are laminated to prevent any direct contact with absorbent material or otherwise product is wrapped in butter paper if region specific restrictions don't allow use of laminated sheets.	
Ultraswachh Disinfection /Sterilization or Gamma Irradiation Sterilization Compatible	Boxes can be sterilized using DRDO's Ultraswachh Sterilization Unit(s) as well as Gamma Irradiation for safety of packaged material for direct utility in OT facility	Gel Technology is compatible for both DRDO's Ultraswachh Sterilization Procedures as well as Gamma Irradiation. User can decide as per available hospital infrastructure.	
	Purchase Information		
Warranty in years(For manufacturing defects)	01 year of comprehensive warranty applicable from Date of Delivery for	Company provides 01 year of comprehensive warranty applicable from	

	any manufacturing defects.	Date of Delivery for any manufacturing Defects as per T&C applied in Manufacturing Warranty Agreement, digitally accessible from Signage Information Sticker (if available as per option selected).
Shelf Life (in years)	Shelf life of item provided should be minimum of 5 years or more.	Shelf Life of item provided in the lot is ≥ 5 years applicable from date of manufacturing.
Residual Shelf Life (in years) – Not Less than ¾ of Shelf Life	Remaining Shelf Life at the time of Supply should be 3/4 th	As per date of manufacturing, the remaining shelf life of provided item is ≥ 3/4 th of applicable Shelf Life of 5 years from Date of Manufacturing.
Certifications & Reports	Copies of reports and certifications to be furnished to buyer on demand	All information i.e. certification/ relevant documentation (reports) in terms of digital copy accessible via QR codes labeled on packs. Specific confidential certifications are provided in physical copies as per requirement furnished by Buyer. It includes GST Certification, MSME Registration with UAN, Vendor Registration with Government Agency (if any), Trademark Registration/ Application (as applicable, if any) and other Company specific documentation requisite for legalized production of items.
Web Authentication of Material Batch	Bath Certification for Supplied Lot should be available online for web based authentication by Buyer.	QR Code with digital access to 'batch certificate' for particular lot Quality Assurance is available outside Item's Pack.
	Sites of Application Info	rmation
Applications (Health Purposes)	Operation Theatres, Intensive Care Units, X-ray, MRI, Special Wards for Comatose/ Semi-Comatose Patients, Rehabilitation Centres for Elderly, Residential Usage for Bed Ridden for Comatose, Semi-Comatose / Non-Comatose Patients.	Item can be deployed in Operation Theatres, ICUs, X-ray diagnostics, MRI, Eye Clinics, Special Wards for both comatose, semicomatose and regular patients, rehabilitation centres for elderly, special children and persons suffering from specific problems, effectively useful in household for all types of bed-riddent, morbid patients including comatose and sem-comatose patients. In addition, health parlors with massages and physiotherapy clinics as well as reception of emergency trauma centers, based on functionality of item, it can be used.

